

TCT

Certificate of Compliance

Certificate No. : TCT200703E907C
Applicant : Medsinglong Global Group Co.,Ltd.
Address : Floor 1st,4th,5th,NO.85th,87th,89th Baiyun Road,Shiqiao Street,Panyu District, Guangzhou city,China.
Manufacturer : Medsinglong Global Group Co.,Ltd.
Address : Floor 1st,4th,5th,NO.85th,87th,89th Baiyun Road,Shiqiao Street,Panyu District, Guangzhou city,China.
Product : Fingertip Pulse Oximeter
Model No. : M170 , M130
Trade mark : N/A

The above products have been tested by us with listed standards and found in compliance with the council EMC 2014/30/EU. It is possible to use CE marking to demonstrate the compliance with this EMC.

Test standards:	Report(s) Number	Issued By	Issued Date
EN 61326-1:2013 EN 61000-3-2:2014 EN 61000-3-3:2013 EN 61326-2-2:2013	TCT200703E907	TCT	Jul. 07, 2020

The statement is based on a single evaluation of one sample of above mentioned products. It does not imply an assessment of the whole production and does not permit the use of the test lab logo.



Tomsin

Tomsin/Senior Engineer

Jul. 07, 2020



Shenzhen TCT Testing Technology Co.,Ltd.

Hotline: 400-6611-140 Tel: 86-755-27673339 Fax: 86-755-27673332 Http://www.tct-lab.com

